

**Certificate of Analysis  
GM501 WASH**

Batchnumber	WASH-1710
Production date	01 September 2017
Expiry date	31 March 2018
Total volume (mL)	15860
Batchnumber HSA	B3AKA00593

	<b>Results</b>	<b>Specifications</b>
pH	7,35	7,20-7,50
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date	19 September 2017
--------------	-------------------

Released by	Vanessa Bierlich
-------------	------------------

Signature

