

**Certificate of Analysis  
GM501 SpermAIR**

Batchnumber	AIR-2002
Production date	12 February 2020
Expiry date	31 August 2020
Batchnumber HSA	B3ALC00833
Total volume (mL)	31860

	<b>Results</b>	<b>Specifications</b>
pH	7,39	7,20-7,50
Osmolality	281	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 27 February 2020

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*