

**Certificate of Analysis
GM501 SpermAIR**

Batchnumber AIR-2403
 Production date 14 March 2024
 Expiry date 30 September 2024
 Batchnumber HSA B03H025323
 Total volume (mL)

Parameter	Results	Specifications
pH	7,42	7,20-7,50
Osmolality	280	270-290 mOsm/kg
Bacterial screen	NO GROWTH	No growth
Fungal screen	NO GROWTH	No growth
Endotoxin	< 0,06	< 0,25 EU/mL
HSA Concentration	4,9	4,0-6,0
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 02 April 2024
 Released by Nicole Geiger
 Signature



REV 08_00 / 2024-01-26

