

## Certificate of Analysis GM501 SpermAIR

Batchnumber	AIR-2411
Production date	06 November 2024
Expiry date	31 May 2025
Batchnumber HSA	B03H100413
Total volume (ml)	31000

Product was sterilized by aseptic processing techniques


Parameter	Results	Specifications
pH	7,37	7,20-7,50
Osmolality	279	270-290 mOsm/kg
Endotoxin level	< 0,06	< 0,25 EU/ml
Concentration HSA	4,8	4,0-6,0 g/l
Sterility test (performed according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
MEA (1-cell to expanded blastocyst within 96 hours)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	21 November 2024
--------------	------------------

Released by	Nicole Geiger
-------------	---------------

Signature



REV 09\_00 / 2024-08-16