

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1701
Production date	25 January 2017
Expiry date	31 July 2017
Total volume (mL)	1500

	Results	Specifications
pH	7,34	7,00-7,60
Osmolality	416	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	09 February 2017
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann