

## Certificate of Analysis GM508 CultActive

Batchnumber	CACT-1702
Production date	08 March 2017
Expiry date	30 September 2017
Total volume (mL)	1500

	Results	Specifications
pH	7,41	7,00-7,60
Osmolality	422	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date	23 March 2017
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*