

Certificate of Analysis GM508 CultActive

Batchnumber	CACT-1705
Production date	04 August 2017
Expiry date	28 February 2018
Total volume (mL)	990

	Results	Specifications
pH	7,29	7,00-7,60
Osmolality	418	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	21 August 2017
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann