

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1805
Production date	06 June 2018
Expiry date	31 December 2018
Total volume (mL)	990

	Results	Specifications
pH	7,32	7,00-7,60
Osmolality	426	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 22 June 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann