

**Certificate of Analysis  
GM508 CultActive**

Batchnumber	CACT-1809
Production date	27 September 2018
Expiry date	31 March 2019
Total volume (mL)	1490

	<b>Results</b>	<b>Specifications</b>
pH	7,26	7,00-7,60
Osmolality	420	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 11 October 2018

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*