

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1903
Production date	01 March 2019
Expiry date	30 September 2019
Total volume (mL)	1000

	Results	Specifications
pH	7,3	7,00-7,60
Osmolality	428	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 18 March 2019

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann