

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1904
Production date	12 April 2019
Expiry date	31 October 2019
Total volume (mL)	1000

	Results	Specifications
pH	7,37	7,00-7,60
Osmolality	425	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 29 April 2019

Released by Vanessa Bierlich

Signature