

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-1911
Production date	14 November 2019
Expiry date	31 May 2020
Total volume (mL)	1000

	Results	Specifications
pH	7,38	7,00-7,60
Osmolality	420	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	04 December 2019
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann