

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-2001
Production date	20 January 2020
Expiry date	31 July 2020
Total volume (mL)	1500

	Results	Specifications
pH	7,28	7,00-7,60
Osmolality	402	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Release date 05 February 2020

Released by Vanessa Bierlich

Signature