

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-2002
Production date	13 February 2020
Expiry date	31 August 2020
Total volume (mL)	1000

	Results	Specifications
pH	7,28	7,00-7,60
Osmolality	423	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	28 February 2020
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann