

**Certificate of Analysis  
GM508 CultActive**

Batchnumber	CACT-2007
Production date	11 August 2020
Expiry date	28 February 2021
Total volume (mL)	1000

	<b>Results</b>	<b>Specifications</b>
pH	7,29	7,00-7,60
Osmolality	424	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Release date	26 August 2020
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*