

**Certificate of Analysis  
GM508 CultActive**

Batchnumber	CACT-2011
Production date	01 December 2020
Expiry date	30 June 2021
Total volume (mL)	1000

	<b>Results</b>	<b>Specifications</b>
pH	7,22	7,00-7,60
Osmolality	425	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 16 December 2020

Released by Vanessa Bierlich

Signature