

Certificate of Analysis GM508 CultActive

Batchnumber	CACT-2109
Production date	27 August 2021
Expiry date	28 February 2022
Total volume (mL)	1000

	Results	Specifications
pH	7,31	7,00-7,60
Osmolality	412	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Release date	13 September 2021
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann