

Certificate of Analysis GM508 CultActive

Batchnumber	CACT-2205
Production date	18 March 2022
Expiry date	30 September 2022
Total volume (mL)	1000

	Results	Specifications
pH	7,42	7,00-7,60
Osmolality	424	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	06 April 2022
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann