

**Certificate of Analysis  
GM508 CultActive**

Batchnumber	CACT-2209
Production date	22 September 2022
Expiry date	31 March 2023
Total volume (mL)	2700

	<b>Results</b>	<b>Specifications</b>
pH	7,35	7,00-7,60
Osmolality	429	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	07 October 2022
Released by	Vanessa Bierlich
Signature	

P.P. 