

Certificate of Analysis GM508 CultActive

Batchnumber	CACT-2210
Production date	18 November 2022
Expiry date	31 May 2023
Total volume (mL)	2700

	Results	Specifications
pH	7,31	7,00-7,60
Osmolality	422	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Release date	05 December 2022
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann