

**Certificate of Analysis  
GM508 CultActive**

Batchnumber	CACT-2211
Production date	29 December 2022
Expiry date	30 June 2023
Total volume (mL)	2000

	<b>Results</b>	<b>Specifications</b>
pH	7,23	7,00-7,60
Osmolality	421	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 13 January 2023

Released by Nicole Geiger

Signature

pp. 

