

**Certificate of Analysis
GM508 CultActive**

Batchnumber	CACT-2304
Production date	07 June 2023
Expiry date	31 December 2023
Total volume (mL)	2000

	Results	Specifications
pH	7,49	7,00-7,60
Osmolality	427	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Release date	22 June 2023
Released by	Nicole Geiger
Signature	

pp. N. Geiger