

**Certificate of Analysis
GM508 CultActive**

Batchnumber CACT-2401
 Production date 23 January 2024
 Expiry date 31 July 2024
 Total volume (ml) 2000

Parameter	Results	Specifications
pH	7,45	7,00-7,60
Osmolality	422	400-450 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/ml
Mouse Embryo Assay (1-cell to blastocyst within 96 hours)	90,48	≥ 80 %

Release date 08 February 2024
 Released by Nicole Geiger
 Signature

