

**Certificate of Analysis  
GM508 CultActive**

Batchnumber CACT-2402  
 Production date 05 February 2024  
 Expiry date 31 August 2024  
 Total volume (ml) 2000

Parameter	Results	Specifications
pH	7,35	7,00-7,60
Osmolality	429	400-450 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/ml
Mouse Embryo Assay (1-cell to blastocyst within 96 hours)	95,24	≥ 80 %

Release date 21 February 2024  
 Released by Nicole Geiger  
 Signature

