

**Certificate of Analysis
GM508 CultActive**

Batchnumber CACT-2410
 Production date 14 November 2024
 Expiry date 31 May 2025
 Total volume (ml) 2700

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,26	7,00-7,60
Osmolality	429	400-450 mOsm/kg
Endotoxin	< 0,05	< 0,25 EU/ml
Sterility test (performed according to the current Ph. Eu. 2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to blastocyst within 96 hours)	100	≥ 80 %

Release date 29 November 2024

Released by Joana Sturm

Signature



