

**Certificate of Analysis
GM508 CultActive**

Batchnumber CACT-2507
 Production date 08 August 2025
 Expiry date 28 February 2026
 Total volume (ml) 2700

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH (@37°C - 6% CO ₂)	7,36	7,00-7,60
Osmolality	426	400-450 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Sterility test (performed according to the current Ph. Eu. 2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to blastocyst within 96 hours)	90	≥ 80 %

Release date 25 August 2025

Released by Joana Sturm

Signature



REV 09_00 / 2025-03-19