

Certificate of Analysis GM508 CultActive

Batchnumber	CACT-2604
Production date	22 April 2026
Expiry date	31 October 2026
Total volume (ml)	2700

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH (@37°C - 6% CO ₂)	7,34	7,00-7,60
Osmolality	427	400-450 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Sterility test (performed according to the current Ph. Eu. 2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to blastocyst within 96 hours)	95	≥ 80 %

Release date	08 May 2026
--------------	-------------

Released by	Nicole Geiger
-------------	---------------

Signature

