

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2006
Production date	16 October 2020
Expiry date	31 July 2021
Total volume (mL)	920

	Results	Specifications
pH	7,38	7,20-7,50
Osmolality	285	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	98	≥ 80 %

Release date	02 November 2020
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann