

Certificate of Analysis
GM501 Cult with Gentamicin and Phenol Red

Batchnumber	CULT-1807PRG
Production date	01 October 2018
Expiry date	30 April 2019
Total volume (mL)	7300
Batchnumber HSA	K814A6871

	Results	Specifications
pH	7,22	7,20-7,45
Osmolality	276	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	16 October 2018
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann