

**Certificate of Analysis
GM501 Cult with Gentamicin**

Batchnumber	CULT-1808G
Production date	12 November 2018
Expiry date	31 May 2019
Total volume (mL)	33700
Batchnumber HSA	K814A6871

	Results	Specifications
pH	7,29	7,20-7,45
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	28 November 2018
Released by	Benthe Brauer

Signature

