

**Certificate of Analysis
GM501 Cult with Gentamicin**

Batchnumber	CULT-1909G
Production date	30 October 2019
Expiry date	30 April 2020
Batchnumber HSA	B3ALD00533
Total volume (mL)	26000

	Results	Specifications
pH	7,25	7,20-7,45
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 14 November 2019

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann