

**Certificate of Analysis
GM501 Cult with Gentamicin**

Batchnumber	CULT-2101G
Production date	10 February 2021
Expiry date	31 August 2021
Batchnumber HSA	K021A6871
Total volume (mL)	13300

	Results	Specifications
pH	7,34	7,20-7,45
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	02 March 2021
Released by	Vanessa Bierlich
Signature	

P.P. 