

Certificate of Analysis
GM501 Cult with Gentamicin and Phenol Red

Batchnumber	CULT-2110PRG
Production date	22 September 2021
Expiry date	31 March 2022
Batchnumber HSA	M106A6671
Total volume (mL)	11800

	Results	Specifications
pH	7,38	7,20-7,45
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 07 October 2021

Released by Vanessa Bierlich

Signature

P.P. 