

**Certificate of Analysis
GM501 Cult with Gentamicin**

Batchnumber	CULT-2204G
Production date	02 June 2022
Expiry date	31 December 2022
Batchnumber HSA	K203D6675
Total volume (mL)	18900

	Results	Specifications
pH	7,26	7,20-7,45
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 22 June 2022

Released by Vanessa Bierlich

Signature