GYNEMED

Certificate of Analysis
GM501 Cult with Gentamicin and Phenol Red

Batchnumber	CULT-2207PRG
Production date	23 September 2022
Expiry date	31 March 2023
Batchnumber HSA	K203D667G
Total volume (mL)	

	Results	**Specifications**
pH	7,31	7,20-7,45
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	86	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	11 October 2022
Released by	Vanessa Bierlich
Signature	

p.p.

GYNEMED GmbH Co. KG, Lübecker Str. 9, 23738 Lensahn, Germany
phone: +49 4363 90329-0, fax: +49 4363 90329-19, e-mail: info@gynemed.de