

Certificate of Analysis
GM501 Cult with Gentamicin and Phenol Red

Batchnumber	CULT-2208PRG
Production date	17 October 2022
Expiry date	30 April 2023
Batchnumber HSA	K203D667G
Total volume (mL)	5000

	Results	Specifications
pH	7,23	7,20-7,45
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 16 November 2022

Released by Nicole Geiger

Signature

