

Certificate of Analysis GM501 Cult with Gentamicin

Batchnumber	CULT-2209G
Production date	18 November 2022
Expiry date	31 May 2023
Batchnumber HSA	K222B6875
Total volume (mL)	15300

	Results	Specifications
pH	7,22	7,20-7,45
Osmolality	275	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	07 December 2022
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann