

Certificate of Analysis
GM501 Cult with Gentamicin

Batchnumber	CULT-2309G
Production date	14 August 2023
Expiry date	29 February 2024
Batchnumber HSA	M306B687D
Total volume (ml)	14500

	Results	Specifications
pH	7,35	7,20-7,45
Osmolality	284	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	30 August 2023
Released by	Nicole Geiger
Signature	