

## Certificate of Analysis GM501 Cult with Gentamicin and Phenol Red

Batchnumber CULT-2309PRG

Production date 14 August 2023

Expiry date 29 February 2024

Batchnumber HSA M306B687D

Total volume (mL) 9600

**Specifications Results** 7,37 7,20-7,45 рН Osmolality 284 270-290 mOsm/kg Bacterial screen **PASSED PASSED** Fungal screen **PASSED PASSED** Endotoxin < 0,05 < 0,25 EU/mL Mouse Embryo Assay (1-cell) 90,48 ≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 30 August 2023

Released by Nicole Geiger

Signature PP. D. Prig

