

## Certificate of Analysis GM501 Cult with Gentamicin and Phenol Red

Batchnumber CULT-2310PRG

Production date 11 September 2023

Expiry date 31 March 2024

Batchnumber HSA M306B687D

Total volume (ml) 7800

Parameter	Results	Specifications
рН	7,43	7,20-7,45
Osmolality	284	270-290 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 05 October 2023

Released by Nicole Geiger

Signature

PP. D. Jeig