

**Certificate of Analysis
GM501 Cult with Gentamicin**

Batchnumber	CULT-2312G
Production date	15 December 2023
Expiry date	30 June 2024
Batchnumber HSA	M306B687M
Total volume (ml)	20000

Parameter	Results	Specifications
pH	7,39	7,20-7,45
Osmolality	279	270-290 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/ml
Mouse Embryo Assay ^(1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	03 January 2024
Released by	Nicole Geiger
Signature	

