

**Certificate of Analysis
GM501 Cult with Gentamicin**

Batchnumber	CULT-2402G
Production date	03 April 2024
Expiry date	31 October 2024
Batchnumber HSA	K331A687A
Total volume (ml)	22300

Parameter	Results	Specifications
pH	7,32	7,20-7,45
Osmolality	277	270-290 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/ml
HSA concentration	9,7	8,0-12,0 mg/ml
Mouse Embryo Assay ^(1-cell)	81	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	18 April 2024
Released by	Joana Sturm

Signature



REV 08_00 / 2024-02-28