

**Certificate of Analysis
GM501 EmbryoStore Freeze**

Batchnumber	ESF-2102
Production date	06 May 2021
Expiry date	30 November 2022
Batchnumber HSA	B3ALE00033
Total volume (kits)	157

	Results	Specifications
pH	7,29	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	25 May 2021
Released by	Vanessa Bierlich

Signature