

**Certificate of Analysis  
GM501 EmbryoStore**

Batchnumber	EST/ESF-1904
Production date	27 August 2019
Expiry date	28 February 2021
Batchnumber HSA	B3ALC00833
Total volume (kits)	25

<b>GM501 EmbryoStore Freeze</b>	<b>Results</b>	<b>Specifications</b>
pH	7,32	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL

<b>GM501 EmbryoStore Thaw1</b>	<b>Results</b>	<b>Specifications</b>
pH	7,32	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL

<b>GM501 EmbryoStore Thaw2</b>	<b>Results</b>	<b>Specifications</b>
pH	7,32	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL

<b>GM501 EmbryoStore Thaw3</b>	<b>Results</b>	<b>Specifications</b>
pH	7,31	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	557	530-560 mOsm/kg

Mouse Embryo Assay (1-cell)	100	≥ 80 %
-----------------------------	-----	--------

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	11 September 2019
--------------	-------------------

Released by	Vanessa Bierlich
-------------	------------------

Signature