

Certificate of Analysis
GM501 EmbryoStore Thaw

Batchnumber	EST-2202
Production date	01 June 2022
Expiry date	31 December 2023
Batchnumber HSA	B3ALF01143
Total volume (kits)	150

GM501 EmbryoStore Thaw1	Results	Specifications
pH	7,28	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
GM501 EmbryoStore Thaw2	Results	Specifications
pH	7,31	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
GM501 EmbryoStore Thaw3	Results	Specifications
pH	7,29	7,20-7,40
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	550	530-560 mOsm/kg
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 22 June 2022

Released by Vanessa Bierlich

Signature

REV 06_00 / 2021-02-26