

**Certificate of Analysis  
GM501 Flush**

Batchnumber	FLUSH-1704
Production date	18 August 2017
Expiry date	28 February 2018
Total volume (mL)	31500

	<b>Results</b>	<b>Specifications</b>
pH	7,36	7,20-7,40
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 04 September 2017

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*