

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-1705
Production date	24 October 2017
Expiry date	30 April 2018
Total volume (mL)	31500

	Results	Specifications
pH	7,26	7,20-7,40
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 08 November 2017

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann