

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-1808
Production date	09 November 2018
Expiry date	31 May 2019
Total volume (mL)	35600

	Results	Specifications
pH	7,3	7,20-7,40
Osmolality	282	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 26 November 2018

Released by Vanessa Bierlich

Signature