

**Certificate of Analysis  
GM501 Flush**

Batchnumber	FLUSH-2002
Production date	10 March 2020
Expiry date	30 September 2020
Total volume (mL)	27000

	<b>Results</b>	<b>Specifications</b>
pH	7,35	7,20-7,40
Osmolality	281	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Release date 25 March 2020

Released by Vanessa Bierlich

Signature