

**Certificate of Analysis
GM501 Flush**

Batchnumber	FLUSH-2004
Production date	02 June 2020
Expiry date	31 December 2020
Total volume (mL)	49000

	Results	Specifications
pH	7,31	7,20-7,40
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date 23 June 2020

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann